



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	RT200RF		
Lot.No.:	202202150101		
Specification:	Volume: <input type="checkbox"/> 20ul <input type="checkbox"/> 50ul <input checked="" type="checkbox"/> 200ul <input type="checkbox"/> 250ul <input type="checkbox"/> 300ul <input type="checkbox"/> 1000ul		
	Filter: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		
	Package Type: <input checked="" type="checkbox"/> rack <input type="checkbox"/> bag <input type="checkbox"/> refill		
Product Name:	Fit LTS pipette tips		
Package:	96pcs/rack, 50racks/case		
Shelf Life:	3 years		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	15th, April, 2024	Report Date	15th, April, 2024
Inspection items and result:			
Cosmetic:	No black spots, no cuts, clear.		
Package:	Accurate quantity and label, inner package without damage		
Dimension:	OK		
Function test:	Accuracy ≥ 98%		
Endotoxin:	<0.03EU/ml		
RNase/DNase:	DNase ≤ 4x10 ⁻⁶ Kunitz units/μl; RNase ≤ 10 ⁻⁹ Kunitz units/μl		
ATP:	2x10 ⁻¹² mg/μl		
PCR Inhibitors:	Not detectable		
Human DNA:	<30pg		
Sterile:	Irradiation, SAL 10 ⁻⁶		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao

PakGent Bioscience (Suzhou) Co., Ltd.
Tel: +86 (0) 512 6670 0017
Address: Floor 1-3, Building 33, Alloy Industrial Park, Wuzhong District, Suzhou, China.

PakGent Bioscience (Shanghai) Co., Ltd.
Tel: 400-167-1717
Address: Building C, No 888, Lake west road, Lin gang section, Free Trade Zone, Shanghai, China.

PakGent Bioscience (Changshu) Co., Ltd.
Tel: +86 (0) 512 5298 8330
Address: No.9 Lecheng Road, Bixi Street, Changshu, Jiangsu Province, China.