



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	WMPB125		
Lot.No.:	202304170101		
Specification:	Size: <input type="checkbox"/> 8ml <input type="checkbox"/> 15ml <input type="checkbox"/> 30ml <input type="checkbox"/> 60ml <input checked="" type="checkbox"/> 125ml <input type="checkbox"/> 250ml <input type="checkbox"/> 500ml <input type="checkbox"/> 1000ml		
	Mouth shape: <input checked="" type="checkbox"/> wide <input type="checkbox"/> narrow color: <input type="checkbox"/> amber <input checked="" type="checkbox"/> natural		
	Bottom material: <input checked="" type="checkbox"/> HDPE <input type="checkbox"/> LDPE <input type="checkbox"/> PP		
Product Name:	Package bottle		
Package:	500pcs/case		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	15th, April, 2024	Report Date	15th, April, 2024
Inspection items and result:			
Cosmetic:	No black spots, no cuts, clear		
Package:	Accurate quantity and label, inner package without damage		
Dimension:	OK		
Function test:	95kpa, 30minutes, no leakage		
Endotoxin:	<0.5EU/ml		
RNase/DNase:	DNase $\leq 4 \times 10^{-6}$ Kunitz units/ μ l; RNase $\leq 10^{-9}$ Kunitz units/ μ l		
PCR Inhibitors:	Not detectable		
Human DNA:	<30pg		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao