



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

<b>Part Number:</b>	PT-5000B-T		
<b>Lot.No.:</b>	202301060101		
<b>Specification:</b>	Volume: <input checked="" type="checkbox"/> 5ml <input type="checkbox"/> 10ml		
	Filter: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No    Package: <input checked="" type="checkbox"/> bag <input type="checkbox"/> rack Applicable brand : <input type="checkbox"/> Eppendorf <input checked="" type="checkbox"/> Thermo		
<b>Product Name:</b>	Pipette tips		
<b>Package:</b>	100pcs/bag, 20bags/case		
<b>Manufacturer:</b>	PakGent Bioscience (Suzhou) Co., Ltd.		
<b>Inspection Date:</b>	15th, April, 2024	<b>Report Date</b>	15th, April, 2024
<b>Inspection items and result:</b>			
<b>Cosmetic:</b>	No black spots, no cuts, clear and transparent.		
<b>Package:</b>	Accurate quantity and label, inner package without damage		
<b>Dimension:</b>	OK		
<b>Function test:</b>	Accuracy $\geq$ 98%		
<b>Endotoxin:</b>	<0.03EU/ml		
<b>RNase/DNase:</b>	DNase $\leq$ 4x10 <sup>-6</sup> Kunitz units/ $\mu$ l; RNase $\leq$ 10 <sup>-9</sup> Kunitz units/ $\mu$ l		
<b>PCR Inhibitors:</b>	Not detectable		
<b>Human DNA:</b>	<30pg		
<b>Conclusion: Comply with Specification</b>			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

**Title: QA Manager**

**Signature: Alvin Cao**