



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	EFG-250P		
Lot.No.:	202403160101		
Specification:	Size: <input type="checkbox"/> 125ml <input checked="" type="checkbox"/> 250ml <input type="checkbox"/> 500ml <input type="checkbox"/> 1000ml		
	Bottom shape: <input checked="" type="checkbox"/> Flat <input type="checkbox"/> Baffled Cap type: <input checked="" type="checkbox"/> Plug cap <input type="checkbox"/> Vent cap		
	Bottle Material: <input checked="" type="checkbox"/> PC <input type="checkbox"/> PETG Cap Material: <input checked="" type="checkbox"/> HDPE		
Product Name:	Erlenmeyer Flask		
Package:	12pcs/case		
Shelf Life:	3 years		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	2th,February,2024	Report Date	2th,February,2024
Inspection items and result:			
Cosmetic:	No black spots, no cuts,clear and transparent		
Package:	Accurate quantity and label, inner package without damage		
Dimension:	OK		
Function test:	N/A		
Endotoxin:	<0.5EU/ml		
RNase/DNase:	DNase $\leq 4 \times 10^{-6}$ Kunitz units/ μ l; RNase $\leq 10^{-9}$ Kunitz units/ μ l		
ATP:	2×10^{-12} mg/ μ l		
PCR Inhibitors:	Not detectable		
Human DNA:	<30pg		
Sterile:	Irradiation,SAL 10^{-6}		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao