



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	CL-PP048		
Lot.No.:	202510180101		
Specification:	Type: <input type="checkbox"/> 6hole <input type="checkbox"/> 12hole <input type="checkbox"/> 24hole <input checked="" type="checkbox"/> 48hole <input type="checkbox"/> 96hole bottom: <input checked="" type="checkbox"/> flat <input type="checkbox"/> U		
Product Name:	Cell culture plate, Poly-D-lysine coated surface treatment		
Package:	Individual pack, 40pcs/case		
Shelf Life:	3 years		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	23th, Oct, 2025	Report Date	23th, Oct, 2025
Inspection items and result:			
Cosmetic:	No black spots, no cuts, clear and transparent		
Package:	Accurate quantity and label, inner package without damage		
Dimension:	Planeness plate ≤ 0.4 mm, lid ≤ 0.4 mm		
Function test:	Cells adhere to the surface and grow in a spindle-shaped, flat or multi-pseudopod form.		
Endotoxin:	< 0.03 EU/ml		
RNase/DNase:	DNase $\leq 4 \times 10^{-6}$ Kunitz units/ μ l; RNase $\leq 10^{-9}$ Kunitz units/ μ l		
ATP:	2×10^{-12} mg/ μ l		
PCR Inhibitors:	Not detectable		
Human DNA:	< 30 pg		
Sterile:	Irradiation, SAL 10^{-6}		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao