



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	CL-MC04		
Lot.No.:	202401220101		
Specification:	Layer: <input type="checkbox"/> 1layer <input type="checkbox"/> 2layers <input checked="" type="checkbox"/> 4layers <input type="checkbox"/> 10layers <input type="checkbox"/> 40layers		
Product Name:	Cell Factory		
Package:	6pcs/case		
Shelf Life:	3 years		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	15th, April, 2024	Report Date	15th, April, 2024
Inspection items and result:			
Cosmetic:	No black spots, no cuts, clear and transparent		
Package:	Accurate quantity and label, inner package without damage		
Dimension:	OK		
Function test:	Comply with specification		
Cloning efficiency with V-79 cell line:	Comply with specification		
Biological Reactivity Tests:	Comply with current edition USP<87>		
ISO10993test:	Comply with current specification ISO10993-1, ISO10993-2, ISO10993-4, ISO10993-11, ISO10993-12		
Endotoxin:	<0.5EU/ml		
Sterile:	Irradiation, SAL10 ⁻⁶		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao