



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	RT1000B-L		
Lot.No. :	202403160101		
Specification:	Volume: <input type="checkbox"/> 20ul <input type="checkbox"/> 50ul <input type="checkbox"/> 200ul <input type="checkbox"/> 250ul <input type="checkbox"/> 300ul <input checked="" type="checkbox"/> 1000ul		
	Filter: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No		
	Package Type: <input type="checkbox"/> rack <input checked="" type="checkbox"/> bag <input type="checkbox"/> refill		
Product Name:	Fit LTS pipette tips(Low Retention)		
Package:	1000pcs/bag, 10bags/case		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	15th, April, 2024	Report Date	15th, April, 2024
Inspection items and result:			
Cosmetic:	No black spots, no cuts, clear.		
Package:	Accurate quantity and label, inner package without damage		
Dimension:	OK		
Function test:	Accuracy \geq 98%		
Endotoxin:	<0.03 EU/ml		
RNase/DNase:	DNase $\leq 4 \times 10^{-6}$ Kunitz units/ μ l; RNase $\leq 10^{-9}$ Kunitz units/ μ l		
PCR Inhibitors:	Not detectable		
Human DNA:	<30 pg		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao