



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

<b>Part Number:</b>	CL-PET048		
<b>Lot.No.:</b>	202510180101		
<b>Specification:</b>	Type: <input type="checkbox"/> 6hole <input type="checkbox"/> 12hole <input type="checkbox"/> 24hole <input checked="" type="checkbox"/> 48hole <input type="checkbox"/> 96hole bottom: <input checked="" type="checkbox"/> flat <input type="checkbox"/> U		
<b>Product Name:</b>	Cell culture plate, TC Enhanced treated		
<b>Package:</b>	Individual pack, 100pcs/case		
<b>Shelf Life:</b>	3 years		
<b>Manufacturer:</b>	PakGent Bioscience (Suzhou) Co., Ltd.		
<b>Inspection Date:</b>	23th, Oct, 2025	<b>Report Date</b>	23th, Oct, 2025
<b>Inspection items and result:</b>			
<b>Cosmetic:</b>	No black spots, no cuts, clear and transparent		
<b>Package:</b>	Accurate quantity and label, inner package without damage		
<b>Dimension:</b>	Planeness plate $\leq 0.4$ mm, lid $\leq 0.4$ mm		
<b>Function test:</b>	Cells adhere to the surface and grow in a spindle-shaped, flat or multi-pseudopod form. contact angle $< 45^\circ$		
<b>Endotoxin:</b>	$< 0.03$ EU/ml		
<b>RNase/DNase:</b>	DNase $\leq 4 \times 10^{-6}$ Kunitz units/ $\mu$ l; RNase $\leq 10^{-9}$ Kunitz units/ $\mu$ l		
<b>ATP:</b>	$2 \times 10^{-12}$ mg/ $\mu$ l		
<b>PCR Inhibitors:</b>	Not detectable		
<b>Human DNA:</b>	$< 30$ pg		
<b>Sterile:</b>	Irradiation, SAL $10^{-6}$		
<b>Conclusion: Comply with Specification</b>			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

**Title: QA Manager**

**Signature: Alvin Cao**