



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

<b>Part Number:</b>	STSS-511CY		
<b>Lot.No.:</b>	202108090101		
<b>Specification:</b>	Volume: <input type="checkbox"/> 0.5ml <input type="checkbox"/> 1.5ml <input checked="" type="checkbox"/> 2.0ml Sterile: <input type="checkbox"/> Y <input checked="" type="checkbox"/> N Tube color: <input checked="" type="checkbox"/> clear <input type="checkbox"/> amber		
	Cap color: <input type="checkbox"/> clear <input type="checkbox"/> red <input type="checkbox"/> blue <input type="checkbox"/> purple <input checked="" type="checkbox"/> yellow <input type="checkbox"/> green <input type="checkbox"/> amber		
	O-ring: <input checked="" type="checkbox"/> Y <input type="checkbox"/> N		
<b>Product Name:</b>	Screw cap tubes		
<b>Package:</b>	500pcs/bag, 10bags/case		
<b>Manufacturer:</b>	PakGent Bioscience (Suzhou) Co., Ltd.		
<b>Inspection Date:</b>	15th, April, 2024	<b>Report Date</b>	15th, April, 2024
<b>Inspection items and result:</b>			
<b>Cosmetic:</b>	No black spots, no cuts, clear and transparent. With correct color identification		
<b>Package:</b>	Accurate quantity and label, inner package without damage		
<b>Dimension:</b>	OK		
<b>Function test:</b>	60kpa test, 15min no leakage		
<b>Endotoxin:</b>	<0.5EU/ml		
<b>RNase/DNase:</b>	DNase $\leq 4 \times 10^{-6}$ Kunitz units/ $\mu$ l; RNase $\leq 10^{-9}$ Kunitz units/ $\mu$ l		
<b>PCR Inhibitors:</b>	Not detectable		
<b>Human DNA:</b>	<30pg		
<b>Conclusion: Comply with Specification</b>			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

**Title: QA Manager**

**Signature: Alvin Cao**