



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

<b>Part Number:</b>	SMB250T		
<b>Lot.No.:</b>	202407040101		
<b>Specification:</b>	Size: <input type="checkbox"/> 125ml <input checked="" type="checkbox"/> 250ml <input type="checkbox"/> 500ml <input type="checkbox"/> 1000ml		
	Bottom shape: <input type="checkbox"/> Flat <input type="checkbox"/> Baffled <input checked="" type="checkbox"/> Square Media		
	Bottle Material: <input checked="" type="checkbox"/> PET <input type="checkbox"/> PETG Cap Material: <input checked="" type="checkbox"/> PP		
<b>Product Name:</b>	Square Media Bottle		
<b>Package:</b>	60pcs/case		
<b>Shelf Life:</b>	3 years		
<b>Manufacturer:</b>	PakGent Bioscience (Suzhou) Co., Ltd.		
<b>Inspection Date:</b>	28th,November,2024	<b>Report Date</b>	28th,November,2024
<b>Inspection items and result:</b>			
<b>Cosmetic:</b>	No black spots, no cuts,clear and transparent		
<b>Package:</b>	Accurate quantity and label, inner package without damage		
<b>Dimension:</b>	OK		
<b>Function test:</b>	Bear-40 °C storage temperature		
<b>Seal test:</b>	95Kpa, 3min no leakage		
<b>Endotoxin:</b>	<0.5EU/ml		
<b>RNase/DNase:</b>	DNase≤4x10 <sup>-6</sup> Kunitz units/μl; RNase≤10 <sup>-9</sup> Kunitz units/μl		
<b>ATP:</b>	2×10 <sup>-12</sup> mg/μl		
<b>PCR Inhibitors:</b>	Not detectable		
<b>Human DNA:</b>	<30pg		
<b>Sterile:</b>	Irradiation,SAL10 <sup>-6</sup>		
<b>Conclusion: Comply with Specification</b>			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

**Title: QA Manager**

**Signature: Alvin Cao**