



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

| | | | |
|----------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------|-------------------|
| Part Number: | SPTB05 | | |
| Lot.No.: | 202402190101 | | |
| Specification: | Volume: <input type="checkbox"/> 1ml <input type="checkbox"/> 2ml <input checked="" type="checkbox"/> 5ml <input type="checkbox"/> 10ml <input type="checkbox"/> 25ml <input type="checkbox"/> 50ml | | |
| | Construction: <input checked="" type="checkbox"/> one piece construction <input type="checkbox"/> durable ultrasonically welded 2 piece | | |
| Product Name: | Serological Pipettes | | |
| Package: | 50pcs/bag, 10bags/case | | |
| Shelf Life: | 4 years | | |
| Manufacturer: | PakGent Bioscience (Suzhou) Co., Ltd. | | |
| Inspection Date: | 15th, April, 2024 | Report Date | 15th, April, 2024 |
| Inspection items and result: | | | |
| Cosmetic: | No black spots, no cuts, clear and transparent. With correct color identification | | |
| Package: | Accurate quantity and label, inner package without damage | | |
| Dimension: | OK | | |
| Function test: | Accuracy $\geq 98\%$ | | |
| Endotoxin: | $< 0.5 \text{ EU/ml}$ | | |
| RNase/DNase: | DNase $\leq 4 \times 10^{-6}$ Kunitz units/ μl ; RNase $\leq 10^{-9}$ Kunitz units/ μl | | |
| ATP: | $2 \times 10^{-12} \text{ mg}/\mu\text{l}$ | | |
| PCR Inhibitors: | Not detectable | | |
| Human DNA: | $< 30 \text{ pg}$ | | |
| Sterile: | Irradiation, SAL 10^{-6} | | |
| Conclusion: Comply with Specification | | | |

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao

PakGent Bioscience (Suzhou) Co., Ltd.

Tel: +86 (0) 512 6670 0017

Address: Floor 1-3, Building 33, Alloy Industrial Park, Wuzhong District, Suzhou, China.

PakGent Bioscience (Shanghai) Co., Ltd.

Tel: 400-167-1717

Address: Building C, No 888, Lake west road, Lin gang section, Free Trade Zone, Shanghai, China.

PakGent Bioscience (Changshu) Co., Ltd.

Tel: +86 (0) 512 5298 8330

Address: No.9 Lecheng Road, Bixi Street, Changshu, Jiangsu Province, China.