



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	STSS-509CY	
Lot.No.:	202311200101	
Specification:	Volume: <input checked="" type="checkbox"/> 0.5ml <input type="checkbox"/> 1.5ml <input type="checkbox"/> 2.0ml Sterile: <input type="checkbox"/> Y <input checked="" type="checkbox"/> N Tube color: <input checked="" type="checkbox"/> clear <input type="checkbox"/> amber Cap color: <input type="checkbox"/> clear <input type="checkbox"/> red <input type="checkbox"/> blue <input type="checkbox"/> purple <input checked="" type="checkbox"/> yellow <input type="checkbox"/> green <input type="checkbox"/> amber O-ring: <input checked="" type="checkbox"/> Y <input type="checkbox"/> N	
	Product Name: Screw cap tubes	
Package:	500pcs/bag, 10bags/case	
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.	
Inspection Date:	2th,February,2024	Report Date 2th,February,2024
Inspection items and result:		
Cosmetic:	No black spots, no cuts,clear and transparent.With correct color identification	
Package:	Accurate quantity and label, inner package without damage	
Dimension:	OK	
Function test:	60kpa test, 15min no leakage	
Endotoxin:	<0.5EU/ml	
RNase/DNase:	DNase $\leq 4 \times 10^{-6}$ Kunitz units/ μ l; RNase $\leq 10^{-9}$ Kunitz units/ μ l	
PCR Inhibitors:	Not detectable	
Human DNA:	<30pg	
Conclusion: Comply with Specification		

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao