



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

| | | |
|--|---|--------------------------------------|
| Part Number: | PCRP-20NS | |
| Lot.No.: | 202312010101 | |
| Specification: | Volume: <input type="checkbox"/> 0.1ml <input checked="" type="checkbox"/> 0.2ml <input type="checkbox"/> 40ul color: <input type="checkbox"/> white <input checked="" type="checkbox"/> clear product: <input checked="" type="checkbox"/> plate <input type="checkbox"/> tube | |
| | Construction: <input checked="" type="checkbox"/> Non skirt <input type="checkbox"/> Half skirt <input type="checkbox"/> High skirt <input type="checkbox"/> Full skirt Printing: <input type="checkbox"/> Y <input checked="" type="checkbox"/> N | |
| Product Name: | PCR consumable | |
| Package: | 10pcs/box, 10boxes/case | |
| Manufacturer: | PakGent Bioscience (Suzhou) Co., Ltd. | |
| Inspection Date: | 25th,October,2024 | Report Date 25th,October,2024 |
| Inspection items and result: | | |
| Cosmetic: | No black spots, no cuts, clear. | |
| Package: | Accurate quantity and label, inner package without damage. | |
| Dimension: | OK | |
| Function test: | Evaporation rate ≤ 10% | |
| Endotoxin: | <0.5EU/ml | |
| RNase/DNase: | DNase ≤ 4x10 ⁻⁶ Kunitz units/μl; RNase ≤ 10 ⁻⁹ Kunitz units/μl | |
| PCR Inhibitors: | Not detectable | |
| Human DNA: | <30pg | |
| Sterile: | No | |
| Conclusion: Comply with Specification | | |

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao