



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	CT50B-MF		
Lot.No.:	202605260101		
Specification:	Volume: <input type="checkbox"/> 0.5ml <input type="checkbox"/> 1.5ml <input type="checkbox"/> 2.0ml <input type="checkbox"/> 5.0ml <input type="checkbox"/> 15ml <input checked="" type="checkbox"/> 50ml		
	Package: <input type="checkbox"/> Rack <input checked="" type="checkbox"/> Bulk Metal Free: <input checked="" type="checkbox"/> Y <input type="checkbox"/> N		
Product Name:	Centrifuge Tube		
Package:	25pcs/bag, 20bags/case		
Shelf Life:	3 years		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	29th,May,2026	Report Date	29th,May,2026
Inspection items and result:			
Cosmetic:	No black spots, no cuts,clear and transparent		
Package:	Accurate quantity and label, inner package without damage		
Dimension:	OK		
Function test:	Centrifuge force 15000xg 10min		
Metal level	<1ppm of common trace metals		
Endotoxin:	<0.03EU/ml		
RNase/DNase:	DNase≤4x10 ⁻⁶ Kunitz units/μl; RNase≤1x10 ⁻⁹ Kunitz units/μl		
ATP:	2×10 ⁻¹² mg/μl		
PCR Inhibitors:	Not detectable		
Human DNA:	<30pg		
Sterile:	Irradiation,SAL10 ⁻⁶		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao