



This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained, for reference.

Part Number:	PD35		
Lot.No.:	202312010101		
	Size: <input checked="" type="checkbox"/> 35mm <input type="checkbox"/> 60mm <input type="checkbox"/> 90mm <input type="checkbox"/> 100mm <input type="checkbox"/> 150mm		
Specification:	Surface treatment: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No		
Product Name:	Petri Dish		
Package:	10pcs/bag, 230bags/case		
Shelf Life:	5 years		
Manufacturer:	PakGent Bioscience (Suzhou) Co., Ltd.		
Inspection Date:	2th,February,2024	Report Date	2th,February,2024
Inspection items and result:			
Cosmetic:	No black spots, no cuts,clear and transparent		
Package:	Accurate quantity and label, inner package without damage		
Dimension:	Planeness Lid ≤0.5mm, Dish ≤0.45mm		
Function test:	N/A		
Endotoxin:	<0.5EU/ml		
RNase/DNase:	DNase ≤4x10 ⁻⁶ Kunitz units/μl; RNase ≤10 ⁻⁹ Kunitz units/μl		
ATP:	2×10 ⁻¹² mg/μl		
PCR Inhibitors:	Not detectable		
Human DNA:	<30pg		
Conclusion: Comply with Specification			

This signature indicates that the above test has met all quality specification and has been reviewed by a quality representative.

Title: QA Manager

Signature: Alvin Cao